

TSP Calibration, Inc.

1110 S. FM 1788 #42
Phone: (432) 296-4992 Email: Chad@tspndt.com
Midland, TX 79706

Certificate of Calibration

Certificate No TSPW-04172026-007

<p>Gage ID PEA-25001 Gage S/N 25001 Description 10 lb Test Weight Operating Procedure: OP-118 Unit of Meas. pounds Manufacturer Parker Cal. Date 4/17/2026 Next Due 4/17/2027 Cal. Freq. 1.00 Years Location Lab</p>	<p>Environmental Conditions</p> <p>Temperature 68 Deg F +/- 2 Humidity 35-55%</p> <p style="text-align: center;">Approved Yes Customer Info. Peak NDT Solutions</p>
--	--

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 15868 Job 15253

As Found: 10.090 lbs.
As Left: 10.090 lbs.

We certify the equipment used for this calibration is traceable to NIST through the following number(s):
Reference Standard: Digital Scale - Serial Number: 8335029565
NIST Numbers: Doc#171842 - Calibration Due Date: 8/21/2026

Calibrated By Chad Adams

Signature



Date: 4/17/2026