

TSP Calibration, Inc.

1110 S. FM 1788 Suite 38 Midland, TX 79706
Phone: (432) 296-4992 Email: Chad@tspndt.com

Certificate of Calibration

Certificate No TSPW-09172025-011

Gage ID PEA-N3560 Gage S/N N3560 Description AC Yoke Operating Procedure: OP-115 Unit of Meas. pounds Manufacturer Magnaflux Cal. Date 9/17/2025 Next Due 3/17/2026 Cal. Freq. 6.00 Months Location Lab	Environmental Conditions Temperature 68 Deg F +/- 2 Humidity 35-55% Approved Yes Customer Info. Peak NDT Solutions
--	---

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 14792 Job 14286

As Found: Passed As Left: Passed

A/C Yoke picked up and held required 10 pound weight for calibration.


We certify the equipment used for this calibration is traceable to NIST through the following number(s):

Reference Standard: Test Weight - Serial Number: 3452

NIST Number: Doc#171848 - Calibration Due Date: 8/21/2026

Calibrated By Chad Adams

Signature

Date: 9/17/2025**APPROVED**

By Kayla Myers at 11:12 am, Sep 18, 2025