

TSP Calibration, Inc.

1110 S. FM 1788 #42
 Phone: (432) 296-4992 Email: Chad@tspndt.com
 Midland, TX 79706

Certificate of Calibration

Certificate No TSPW-03162026-006

| | |
|--|--|
| <p>Gage ID PEA-1145729 Gage S/N 1145729 Description 12" Digital Caliper Operating Procedure: OP-110 Unit of Meas. inch Manufacturer Mitutoyo Cal. Date 3/16/2026 Next Due 9/16/2026 Cal. Freq. 6.00 Months Location Lab</p> | <p>Environmental Conditions Temperature 68 Deg F +/- 2 Humidity 35-55%</p> <p style="text-align: center;">Approved Yes Customer Info. Peak NDT Solutions</p> |
|--|--|

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 15710 Job 15208

| Standard ID | 1.000 " | Uncertainty | Units | | Type V | |
|---------------------|------------------|-------------|--------|----------|--------|----------------|
| Limited Use? | No | Minimum | 0.999 | Nominal | 1.000 | Maximum 1.001 |
| Ref Type | 81PC Gage Blocks | As Found | 1.000 | Accuracy | 0.000 | Fail Before No |
| | | As Left | 1.000 | Accuracy | 0.000 | Fail After No |
| Gage ID of Standard | REF-070111 | | | | | |
| Std Due Date | 1/26/2027 | Gage S/N | 070111 | NIST No. | 81004 | |

| Standard ID | 3.000 " | Uncertainty | Units | | Type V | |
|---------------------|------------------|-------------|--------|----------|--------|----------------|
| Limited Use? | No | Minimum | 2.999 | Nominal | 3.000 | Maximum 3.001 |
| Ref Type | 81PC Gage Blocks | As Found | 3.000 | Accuracy | 0.000 | Fail Before No |
| | | As Left | 3.000 | Accuracy | 0.000 | Fail After No |
| Gage ID of Standard | REF-070111 | | | | | |
| Std Due Date | 1/26/2027 | Gage S/N | 070111 | NIST No. | 81004 | |

| Standard ID | 5.000 " | Uncertainty | Units | | Type V | |
|--------------|---------|-------------|-------|---------|--------|---------------|
| Limited Use? | No | Minimum | 4.999 | Nominal | 5.000 | Maximum 5.001 |

| | | | | | | | |
|----------|------------------|----------|-------|----------|-------|-------------|----|
| Ref Type | 81PC Gage Blocks | As Found | 5.000 | Accuracy | 0.000 | Fail Before | No |
| | | As Left | 5.000 | Accuracy | 0.000 | Fail After | No |

Gage ID of Standard REF-070111

Std Due Date 1/26/2027

Gage S/N 070111

NIST No. 81004

| | | | | | | | |
|--------------|------------------|-------------|-------|----------|-------|-------------|-------|
| Standard ID | 7.000 " | Uncertainty | | Units | | Type | V |
| Limited Use? | No | Minimum | 6.999 | Nominal | 7.000 | Maximum | 7.001 |
| Ref Type | 81PC Gage Blocks | As Found | 7.000 | Accuracy | 0.000 | Fail Before | No |
| | | As Left | 7.000 | Accuracy | 0.000 | Fail After | No |

Gage ID of Standard REF-070111

Std Due Date 1/26/2027

Gage S/N 070111

NIST No. 81004

| | | | | | | | |
|--------------|------------------|-------------|-------|----------|-------|-------------|-------|
| Standard ID | 9.000 " | Uncertainty | | Units | | Type | V |
| Limited Use? | No | Minimum | 8.999 | Nominal | 9.000 | Maximum | 9.001 |
| Ref Type | 81PC Gage Blocks | As Found | 9.000 | Accuracy | 0.000 | Fail Before | No |
| | | As Left | 9.000 | Accuracy | 0.000 | Fail After | No |

Gage ID of Standard REF-070111

Std Due Date 1/26/2027

Gage S/N 070111

NIST No. 81004

Calibrated By Chad Adams Signature  Date: 3/16/2026

APPROVED
By Kayla Myers at 2:45 pm, Mar 20, 2026