

TSP Calibration, Inc.

1110 S. FM 1788 Suite 38 Midland, TX 79706
Phone: (432) 296-4992 Email: Chad@tspndt.com

Certificate of Calibration

Certificate No TSPW-03192025-006

Gage ID PEA-1234 Gage S/N 1234 Description 20-0-20 gauss meter Operating Procedure: OP-117 Unit of Meas. gauss Manufacturer R.B. Annis Cal. Date 3/19/2025 Next Due 9/19/2025 Cal. Freq. 6.00 Months Location Lab	Environmental Conditions Temperature 68 Deg F +/- 2 Humidity 35-55% Approved Yes Customer Info. Peak NDT Solutions
--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------	---------------------------------------------------------------------------------------------------------------------------------------------------------

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 14050 Job 13175

Standard ID	+10	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	9	Nominal	10	Maximum	11
Ref Type	20-0-20 Gauss Meter	As Found	10	Accuracy	0	Fail Before	No
		As Left	10	Accuracy	0	Fail After	No
Gage ID of Standard	REF-TSPSTD-02						
Std Due Date	8/8/2025	Gage S/N	TSPSTD-02	NIST No.	154784		
Standard ID	+20	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	19	Nominal	20	Maximum	21
Ref Type	20-0-20 Gauss Meter	As Found	20	Accuracy	0	Fail Before	No
		As Left	20	Accuracy	0	Fail After	No
Gage ID of Standard	REF-TSPSTD-02						
Std Due Date	8/8/2025	Gage S/N	TSPSTD-02	NIST No.	154784		
Standard ID	-10	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	9	Nominal	10	Maximum	11

Ref Type	20-0-20 Gauss Meter	As Found	10	Accuracy	0	Fail Before	No
		As Left	10	Accuracy	0	Fail After	No

Gage ID of Standard REF-TSPSTD-02

Std Due Date 8/8/2025

Gage S/N TSPSTD-02 NIST No. 154784

Standard ID	-20	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	19	Nominal	20	Maximum	21
Ref Type	20-0-20 Gauss Meter	As Found	20	Accuracy	0	Fail Before	No
		As Left	20	Accuracy	0	Fail After	No


Gage ID of Standard REF-TSPSTD-02

Std Due Date 8/8/2025

Gage S/N TSPSTD-02 NIST No. 154784

Calibrated By Chad Adams

Signature



Date: 3/19/2025

APPROVED

By Kayla Myers at 5:13 pm, Mar 24, 2025