

## TSP Calibration, Inc.

1110 S. FM 1788 #42  
 Phone: (432) 296-4992 Email: Chad@tspndt.com  
 Midland, TX 79706

### Certificate of Calibration

Certificate No TSPW-03132026-024

<p><b>Gage ID</b> PEA-230217782  <b>Gage S/N</b> 230217782  <b>Description</b> 0-1" Groove Micrometer  <b>Operating Procedure:</b> OP-111  <b>Unit of Meas.</b> Inch  <b>Manufacturer</b> Insize  <b>Cal. Date</b> 3/13/2026  <b>Next Due</b> 9/13/2026  <b>Cal. Freq.</b> 6.00 Months  <b>Location</b> Lab</p>	<p><b>Environmental Conditions</b>                  Temperature 68 Deg F +/- 2                  Humidity 35-55%</p> <p style="text-align: center;"><b>Approved</b> Yes  <b>Customer Info.</b> Peak NDT Solutions</p>
---	--

**Certification Statement**

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

**Findings**

Master	Actual	Deviation
0.100	0.100	0.000
0.300	0.300	0.000
0.500	0.500	0.000
0.700	0.700	0.000
0.900	0.900	0.000

APPROVED


By Kayla Myers at 2:16 pm, May 28, 2026

Tol +0.001 / -0.001

NOTE: IF ONLY AN "AS FOUND" READING IS STATED, "AS FOUND" AND "AS LEFT" READINGS ARE THE SAME WITH NO ADJUSTMENTS.

We certify the equipment used for this calibration is traceable to NIST through the following number(s):  
 Reference Standard: Gauge Blocks - Serial Number: 070111  
 NIST Number: 81004 - Calibration Due Date: 1/26/2027

---

Calibrated By Chad Adams      Signature       Date: 3/13/2026