

**TSP Calibration, Inc.**

1110 S. FM 1788 Suite 38 Midland, TX 79706  
Phone: (432) 296-4992 Email: Chad@tspndt.com

**Certificate of Calibration**

Certificate No TSPW-01062025-001

<b>Gage ID</b> PEA-230721047 <b>Gage S/N</b> 230721047 <b>Description</b> 0-1" Groove Micrometer <b>Operating Procedure:</b> OP-111 <b>Unit of Meas.</b> inch <b>Manufacturer</b> insize <b>Cal. Date</b> 1/6/2025 <b>Next Due</b> 7/6/2025 <b>Cal. Freq.</b> 6.00 Months <b>Location</b> Lab	<b>Environmental Conditions</b>  Temperature 68 Deg F +/- 2 Humidity 35-55%  <b>Approved</b> Yes <b>Customer Info.</b> Peak NDT Solutions
--	---

**Certification Statement**

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

**Findings**

SO 13719 Job 11839

0-1" Groove Micrometer Measured Values As Found

Master	Actual	Deviation
+0.100	+0.100	+0.000
+0.300	+0.300	+0.000
+0.500	+0.500	+0.000
+0.700	+0.700	+0.000
+0.900	+0.900	+0.000

Tol +0.001 / -0.001

NOTE: IF ONLY AN "AS FOUND" READING IS STATED, "AS FOUND AND AS LEFT" READINGS ARE THE SAME WITH NO ADJUSTMENTS.

We certify the equipment used for this calibration is traceable to NIST through one or more of the following numbers:

Reference Standard Serial #: 070111  
GAGE BLOCKS  
NIST Number: 81004  
Cal Date: 1/26/2024 Due Date: 1/26/2027

Calibrated By Chad Adams

Signature 

Date: 1/6/2025

**APPROVED**  
*By Kayla Myers at 4:29 pm, Jan 07, 2025*