

TSP Calibration, Inc.

1110 S. FM 1788 Suite 38 Midland, TX 79706
Phone: (432) 296-4992 Email: Chad@tspndt.com

Certificate of Calibration

Certificate No TSPW-09152025-005

| | |
|--|---|
| Gage ID PEA-230217782 Gage S/N 230217782 Description 0-1" Groove Micrometer Operating Procedure: OP-111 Unit of Meas. inch Manufacturer insize Cal. Date 9/15/2025 Next Due 3/15/2026 Cal. Freq. 6.00 Months Location Lab | Environmental Conditions Temperature 68 Deg F +/- 2 Humidity 35-55% Approved Yes Customer Info. Peak NDT Solutions |
|--|---|

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

0-1" Groove Micrometer Measured Values As Found

| Master | Actual | Deviation |
|--------|--------|-----------|
| 0.100 | 0.100 | 0.000 |
| 0.300 | 0.300 | 0.000 |
| 0.500 | 0.500 | 0.000 |
| 0.700 | 0.700 | 0.000 |
| 0.900 | 0.900 | 0.000 |

APPROVED**By Kayla Myers at 10:09 am, Sep 23, 2025**

Tol +0.001 / -0.001

NOTE: IF ONLY AN "AS FOUND" READING IS STATED, "AS FOUND" AND "AS LEFT" READINGS ARE THE SAME WITH NO ADJUSTMENTS.

We certify the equipment used for this calibration is traceable to NIST through the following number(s):
Reference Standard: Gauge Blocks - Serial Number: 070111
NIST Number: 81004 - Calibration Due Date: 1/26/2027

Calibrated By Chad Adams

Signature

Date: 9/15/2025