

TSP Calibration, Inc.

1110 S. FM 1788 #42
Phone: (432) 296-4992 Email: Chad@tspndt.com
Midland, TX 79706

Certificate of Calibration

Certificate No TSPW-10152025-041

Gage ID PEA-895 Gage S/N 895 Description Magnetic Particle Unit Operating Procedure: OP-120 Unit of Meas. Amps Manufacturer Parker Cal. Date 10/15/2025 Next Due 4/15/2026 Cal. Freq. 6.00 Months Location Lab	Environmental Conditions Temperature 68 Deg F +/- 2 Humidity 35-55% Approved Yes Customer Info. Peak NDT Solutions
---	---

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 14905 Job 11120

Shot (Amp) as Found

Master	Actual	Deviation
1040	1000	+60
1010	975	+35
910	870	+40

Tolerance: +/- 10%

We certify the equipment used for this calibration is traceable to NIST through the following number(s):

Reference Standard: ShotChek II - Serial Number: 5100

NIST Number: 1112318 & 1112316 - Calibration Due Date: 10/22/2025

Calibrated By Chad AdamsSignature Date: 10/15/2025**APPROVED**

By Kayla Myers at 4:26 pm, Nov 11, 2025