

TSP Calibration, Inc.

1110 S. FM 1788 Suite 38 Midland, TX 79706
Phone: (432) 296-4992 Email: Chad@tspndt.com

Certificate of Calibration

Certificate No TSPW-03192025-007

Gage ID PEA-PK908 Gage S/N PK908 Description 10 1/2" Magnetic Coil Operating Procedure: OP-116 Unit of Meas. gauss Manufacturer Cut-N-Edge Cal. Date 3/19/2025 Next Due 9/19/2025 Cal. Freq. 6.00 Months Location Lab	Environmental Conditions Temperature 68 Deg F +/- 2 Humidity 35-55% Approved Yes Customer Info. Peak NDT Solutions
--	---

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 14050 Job 13175

AS FOUND: PASSED AS LEFT: PASSED

Mag coil putting out magnetic field from both the pos. and neg. sides of coil equal to or greater than 50 gauss for calibration.

Calibrated 50-0-50 gauss meter serial no. TSPSTD-01 was used for this calibration.

Calibration due date for gauss meter is on 8/8/2025

Gauss meter is traceable to NIST by the following number: Certification #154783

Calibrated By Chad Adams

Signature

Date: 3/19/2025**APPROVED****By Kayla Myers at 5:14 pm, Mar 24, 2025**