



Tubular Inspection Products

1551 Windway
Odessa, TX 79763
(432) 363-2200

PO Box 862
Odessa, TX 79760
Fax: (432) 363-2202

CERTIFICATE OF CALIBRATION

Traceable to NIST — ISO/IEC 17025:2017

This Calibration Certificate confirms that the instruments described herein have been calibrated in compliance with the requirements set forth by ISO/IEC 17025:2017, ensuring that all measurement results are traceable to national or international standards, including those established by the National Institute of Standards and Technology (NIST). This certificate reflects adherence to the manufacturer's specifications, demonstrating compliance with established guidelines for performance and reliability.

Standards: ISO 17025:2017 • ISO 9001 • API Q1/7-2

INSTRUMENT & CALIBRATION DETAILS

Customer	Peak NDT Solutions, LLC	Manufacturer	N/A
Calibration Date	September 21, 2026	TIP Procedure	TIP-COILW01.1
Calibration Due Date	March 21, 2027	Calibration Technician	Freddy Seabourn
Serial Number	PK-908	Order Number	81714
Instrument	10 1/2 AC Rectified Coil		

CALIBRATION RESULTS

Polarity Check	Gauss Reading	Result
Positive	209 G	PASS
Zero	0 G	—
Negative	-208 G	PASS

INSTRUMENT STATUS SUMMARY

Instrument Received	Instrument Returned	Instrument Status
In Tolerance	In Tolerance	PASSED

CALIBRATION STANDARDS USED (TRACEABILITY)

- TIP Master Gauss Meter, Alpha Lab Inc. GM-1 Digital Gauss Meter, S/N 3843 (Cert. Due 01/28/2027)
- TIP Master Hygrometer, S/N 31874 / 1H2048MCR / 130070751 (Cert. Due 01/28/2027)

Calibration Technician

Freddy Seabourn

Issue Date: September 23, 2026

APPROVED

By Kayla Myers at 1:07 pm, Sep 24, 2026