Gage ID: PEA-PK908

## TSP Calibration, Inc.

1110 S. FM 1788 Suite 38 Midland, TX 79706 Phone: (432) 296-4992 Email: Chad@tspndt.com

## **Certificate of Calibration**

#### Certificate No TSPW-09172025-010

Gage ID PEA-PK908

Gage S/N PK908

Description 10 1/2" Magnetic Coil

Operating OP-116

Procedure:

Unit of Meas. gauss

Manufacturer Cut-N-Edge

Cal. Date 9/17/2025 Next Due 3/17/2026

Cal. Freq. 6.00

Months

Location Lab

**Environmental Conditions** 

Temperature

68 Deg F +/- 2

Humidity

35-55%

Approved Yes

Customer Info. Peak NDT Solutions

#### **Certification Statement**

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

### **Findings**

SO 14792 Job 14286

As Found: Passed

As Left: Passed

Coil putting out magnetic field from both the positive and negative side equal to or greater than 50 gauss for calibration.

We certify the equipment used for this calibration is traceable to NIST through the following number(s):

Reference Standard: 50-0-50 Magnetometer - Serial Number: TSPSTD-01

NIST Number: Doc#171843 - Calibration Due Date: 8/21/2026

Calibrated By **Chad Adams**  Signature

Date:

9/17/2025

# **APPROVED**

By Kayla Myers at 11:11 am, Sep 18, 2025