

TSP Calibration, Inc.

1110 S. FM 1788 #42
Phone: (432) 296-4992 Email: Chad@tspndt.com
Midland, TX 79706

Certificate of Calibration

Certificate No TSPW-03172026-013

<p>Gage ID PEA-PK908 Gage S/N PK908 Description 10 1/2" Magnetic Coil Operating Procedure: OP-116 Unit of Meas. gauss Manufacturer Cut-N-Edge Cal. Date 3/17/2026 Next Due 9/17/2026 Cal. Freq. 6.00 Months Location Lab</p>	<p>Environmental Conditions Temperature 68 Deg F +/- 2 Humidity 35-55%</p> <p style="text-align: center;">Approved Yes Customer Info. Peak NDT Solutions</p>
---	--

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 15710 Job 15208

As Found: Passed As Left: Passed

A/C Yoke picked up and held required 10 pound weight for calibration.

We certify the equipment used for this calibration is traceable to NIST through the following number(s):

Reference Standard: Test Weight - Serial Number: 3452
NIST Number: Doc#171848 - Calibration Due Date: 8/21/2026

Calibrated By Chad Adams Signature  Date: 3/17/2026

APPROVED
By Kayla Myers at 6:26 pm, Mar 17, 2026