

TSP Calibration, Inc.

1110 S. FM 1788 #42
Phone: (432) 296-4992 Email: Chad@tspndt.com
Midland, TX 79706

Certificate of Calibration

Certificate No TSPW-04202026-000

<p>Gage ID PEA-895 Gage S/N 895 Description Magnetic Particle Unit Operating Procedure: OP-120 Unit of Meas. Amps Manufacturer Parker Cal. Date 4/20/2026 Next Due 10/20/2026 Cal. Freq. 6.00 Months Location Lab</p>	<p>Environmental Conditions Temperature 68 Deg F +/- 2 Humidity 35-55%</p> <p style="text-align: center;">Approved Yes Customer Info. Peak NDT Solutions</p>
--	--

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 15891 Job 15253

Shot (Amp) as Found

Master	Actual	Deviation
800	840	+40
800	840	+40
800	830	+30

APPROVED
By Kayla Myers at 3:56 pm, May 19, 2026

Tolerance: +/- 10%

We certify the equipment used for this calibration is traceable to NIST through the following number(s):
ShotChek II, Indutronics, Model II - Serial Number 5100
NIST Number 1112318 & 1112316 - Calibration Due Date: 10/27/2026

Calibrated By Chad Adams Signature  Date: 4/20/2026