

TSP Calibration, Inc.

1110 S. FM 1788 Suite 38 Midland, TX 79706
 Phone: (432) 296-4992 Email: Chad@tspndt.com

Certificate of Calibration

Certificate No TSPW-07082025-015

<p>Gage ID PEA-JZ29JF0009 Gage S/N JZ29JF0009 Description Pit Gauge Operating Procedure: OP-122 Unit of Meas. inch Manufacturer Gagemaker Cal. Date 7/8/2025 Next Due 7/8/2026 Cal. Freq. 1.00 Years Location Lab</p>	<p>Environmental Conditions</p> <p>Temperature 68 Deg F +/- 2 Humidity 35-55%</p> <p style="text-align: center;">Approved Yes Customer Info. Peak NDT Solutions</p>
-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------	---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 14534 Job 14212

Standard ID .100 "	Uncertainty		Units		Type V
Limited Use? No	Minimum	0.099	Nominal	0.100	Maximum 0.101
Ref Type Bench Calibrator	As Found	0.100	Accuracy	0.000	Fail Before No
	As Left	0.100	Accuracy	0.000	Fail After No

Gage ID of Standard REF-I132869
Std Due Date 7/26/2025 **Gage S/N** I132869 **NIST No.** 138462

Standard ID .200 "	Uncertainty		Units		Type V
Limited Use? No	Minimum	0.199	Nominal	0.200	Maximum 0.201
Ref Type Bench Calibrator	As Found	0.200	Accuracy	0.000	Fail Before No
	As Left	0.200	Accuracy	0.000	Fail After No

Gage ID of Standard REF-I132869
Std Due Date 7/26/2025 **Gage S/N** I132869 **NIST No.** 138462

Standard ID .300 "	Uncertainty		Units		Type V
Limited Use? No	Minimum	0.299	Nominal	0.300	Maximum 0.301

Ref Type	Bench Calibrator	As Found	0.300	Accuracy	0.000	Fail Before	No
		As Left	0.300	Accuracy	0.000	Fail After	No

Gage ID of Standard	REF-I132869	Gage S/N	I132869	NIST No.	138462
Std Due Date	7/26/2025				

Standard ID	.400 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.399	Nominal	0.400	Maximum	0.401
Ref Type	Bench Calibrator	As Found	0.400	Accuracy	0.000	Fail Before	No
		As Left	0.400	Accuracy	0.000	Fail After	No

Gage ID of Standard	REF-I132869	Gage S/N	I132869	NIST No.	138462
Std Due Date	7/26/2025				

Standard ID	.500 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.499	Nominal	0.500	Maximum	0.501
Ref Type	Bench Calibrator	As Found	0.500	Accuracy	0.000	Fail Before	No
		As Left	0.500	Accuracy	0.000	Fail After	No

Gage ID of Standard	REF-I132869	Gage S/N	I132869	NIST No.	138462
Std Due Date	7/26/2025				

Calibrated By Chad Adams Signature  Date: 7/8/2025

APPROVED
By Kayla Myers at 1:33 pm, Jul 30, 2025