

## TSP Calibration, Inc.

1110 S. FM 1788 Suite 38 Midland, TX 79706  
Phone: (432) 296-4992 Email: Chad@tspndt.com

### Certificate of Calibration

Certificate No TSPW-09182024-008

<b>Gage ID</b> PEA-J02116 <b>Gage S/N</b> J02116 <b>Description</b> 5-step block <b>Operating Procedure:</b> OP-126 <b>Unit of Meas.</b> inch <b>Manufacturer</b> N/A <b>Cal. Date</b> 9/18/2024 <b>Next Due</b> 9/18/2025 <b>Cal. Freq.</b> 1.00      Years <b>Location</b> Lab	<b>Environmental Conditions</b> Temperature 68 Deg F +/- 2 Humidity 35-55%  <b>Approved</b> Yes <b>Customer Info.</b> Peak NDT Solutions
---	---

#### Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

#### Findings

SO 13306 Job 13203

<b>Standard ID</b>	.100 "	<b>Uncertainty</b>		<b>Units</b>		<b>Type</b>	V
<b>Limited Use?</b>	No	<b>Minimum</b>	0.098	<b>Nominal</b>	0.100	<b>Maximum</b>	0.102
<b>Ref Type</b>	12" Digital Caliper	<b>As Found</b>	0.100	<b>Accuracy</b>	0.000	<b>Fail Before</b>	No
		<b>As Left</b>	0.100	<b>Accuracy</b>	0.000	<b>Fail After</b>	No
<b>Gage ID of Standard</b>	REF-1095377						
<b>Std Due Date</b>	8/8/2025	<b>Gage S/N</b>	1095377	<b>NIST No.</b>	154785		
<b>Standard ID</b>	.200 "	<b>Uncertainty</b>		<b>Units</b>		<b>Type</b>	V
<b>Limited Use?</b>	No	<b>Minimum</b>	0.198	<b>Nominal</b>	0.200	<b>Maximum</b>	0.202
<b>Ref Type</b>	12" Digital Caliper	<b>As Found</b>	0.200	<b>Accuracy</b>	0.000	<b>Fail Before</b>	No
		<b>As Left</b>	0.200	<b>Accuracy</b>	0.000	<b>Fail After</b>	No
<b>Gage ID of Standard</b>	REF-1095377						
<b>Std Due Date</b>	8/8/2025	<b>Gage S/N</b>	1095377	<b>NIST No.</b>	154785		
<b>Standard ID</b>	.300 "	<b>Uncertainty</b>		<b>Units</b>		<b>Type</b>	V
<b>Limited Use?</b>	No	<b>Minimum</b>	0.298	<b>Nominal</b>	0.300	<b>Maximum</b>	0.302

Ref Type	12" Digital Caliper	As Found	0.300	Accuracy	0.000	Fail Before	No
		As Left	0.300	Accuracy	0.000	Fail After	No

Gage ID of Standard REF-1095377

Std Due Date 8/8/2025

Gage S/N 1095377 NIST No. 154785

Standard ID	.400 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.398	Nominal	0.400	Maximum	0.402
Ref Type	12" Digital Caliper	As Found	0.400	Accuracy	0.000	Fail Before	No
		As Left	0.400	Accuracy	0.000	Fail After	No

Gage ID of Standard REF-1095377

Std Due Date 8/8/2025

Gage S/N 1095377 NIST No. 154785

Standard ID	.500 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.498	Nominal	0.500	Maximum	0.502
Ref Type	12" Digital Caliper	As Found	0.500	Accuracy	0.000	Fail Before	No
		As Left	0.500	Accuracy	0.000	Fail After	No

Gage ID of Standard REF-1095377

Std Due Date 8/8/2025

Gage S/N 1095377 NIST No. 154785

Calibrated By Chad Adams

Signature



Date: 9/18/2024

**APPROVED**

By Kayla Myers at 3:16 pm, Sep 24, 2024