

TSP Calibration, Inc.

1110 S. FM 1788 Suite 38 Midland, TX 79706
Phone: (432) 296-4992 Email: Chad@tspndt.com

Certificate of Calibration

Certificate No TSPW-01202025-010

Gage ID PEA-25001 Gage S/N 25001 Description 10 lb Test Weight Operating Procedure: OP-118 Unit of Meas. pounds Manufacturer Parker Cal. Date 1/20/2025 Next Due 1/20/2026 Cal. Freq. 1.00 Years Location Lab	Environmental Conditions Temperature 68 Deg F +/- 2 Humidity 35-55% Approved Yes Customer Info. Peak NDT Solutions
--	---

Certification Statement


TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

AS FOUND: PASSED AS LEFT: PASSED

Using calibrated scale serial no. 8335029565
Test weight as found 10.090 lbs.
Test weight as left 10.090 lbs.

Calibration due date for scale is on 8/8/2025
Scale is traceable to NIST by the following numbers: Doc# 154780

Calibrated By Chad AdamsSignature Date: 1/20/2025**APPROVED****By Kayla Myers at 3:18 pm, Jan 22, 2025**