

TSP Calibration, Inc.

1110 S. FM 1788 #42
 Phone: (432) 296-4992 Email: Chad@tspndt.com
 Midland, TX 79706

Certificate of Calibration

Certificate No TSPW-03172026-012

<p>Gage ID PEA-46552 Gage S/N 46552 Description UT Thickness Gauge Operating Procedure: OP-121 Unit of Meas. Inch Manufacturer Dakota Cal. Date 3/17/2026 Next Due 9/17/2026 Cal. Freq. 6.00 Months Location Lab</p>	<p>Environmental Conditions Temperature 68 Deg F +/- 2 Humidity 35-55%</p> <p style="text-align: center;">Approved Yes Customer Info. Peak NDT Solutions</p>
---	--

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 15710 Job 15208

Standard ID	Uncertainty	Units	Type
.100 "			V
Limited Use?	Minimum	0.098 Nominal	Maximum 0.102
Ref Type	As Found	0.100 Accuracy	Fail Before No
	As Left	0.100 Accuracy	Fail After No
Gage ID of Standard REF-337214			
Std Due Date 8/21/2026		Gage S/N 337214 NIST No. 171846	

Standard ID	Uncertainty	Units	Type
.200 "			V
Limited Use?	Minimum	0.198 Nominal	Maximum 0.202
Ref Type	As Found	0.200 Accuracy	Fail Before No
	As Left	0.200 Accuracy	Fail After No
Gage ID of Standard REF-337214			
Std Due Date 8/21/2026		Gage S/N 337214 NIST No. 171846	

Standard ID	Uncertainty	Units	Type
.300 "			V
Limited Use?	Minimum	0.298 Nominal	Maximum 0.302

Ref Type	5 Step Block	As Found	0.300	Accuracy	0.000	Fail Before	No
		As Left	0.300	Accuracy	0.000	Fail After	No

Gage ID of Standard	REF-337214	Gage S/N	337214	NIST No.	171846
Std Due Date	8/21/2026				

Standard ID	.400 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.398	Nominal	0.400	Maximum	0.402
Ref Type	5 Step Block	As Found	0.400	Accuracy	0.000	Fail Before	No
		As Left	0.400	Accuracy	0.000	Fail After	No

Gage ID of Standard	REF-337214	Gage S/N	337214	NIST No.	171846
Std Due Date	8/21/2026				

Standard ID	.500 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.498	Nominal	0.500	Maximum	0.502
Ref Type	5 Step Block	As Found	0.500	Accuracy	0.000	Fail Before	No
		As Left	0.500	Accuracy	0.000	Fail After	No

Gage ID of Standard	REF-337214	Gage S/N	337214	NIST No.	171846
Std Due Date	8/21/2026				

Calibrated By Chad Adams Signature  Date: 3/17/2026

APPROVED
By Kayla Myers at 6:25 pm, Mar 17, 2026