

## TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518  
 Phone: (337) 236-6078 Email: taichi@tspndt.com

### Certificate of Calibration

Certificate No TSP-01122026-009

<p><b>Gage ID</b> PEA-BR10LL0050  <b>Gage S/N</b> BR10LL0050  <b>Description</b> Lead Gauge  <b>Operating Procedure:</b> OP-124  <b>Unit of Meas.</b> inch  <b>Manufacturer</b> Gagemaker  <b>Cal. Date</b> 1/12/2026  <b>Next Due</b> 7/12/2026  <b>Cal. Freq.</b> 6.00                      Months  <b>Location</b> Lab</p>	<p><b>Environmental Conditions</b>                  Temperature 68 +/- 2 deg F                  Humidity 20-55%</p> <p style="text-align: center;"><b>Approved Yes</b>  <b>Customer Info.</b> Peak NDT</p>
---	--

**Certification Statement**

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

**Findings**

<b>Standard ID</b>	+0.00500"	<b>Uncertainty</b>		<b>Units</b>		<b>Type</b>	V
<b>Limited Use?</b>	No	<b>Minimum</b>	0.00450	<b>Nominal</b>	0.00500	<b>Maximum</b>	0.00550
<b>Ref Type</b>	Mic Trac	<b>As Found</b>	0.00491	<b>Accuracy</b>	-0.00009	<b>Fail Before</b>	No
		<b>As Left</b>	0.00491	<b>Accuracy</b>	-0.00009	<b>Fail After</b>	No
<b>Gage ID of Standard</b>	REF-G141308						
<b>Std Due Date</b>	10/1/2026	<b>Gage S/N</b>	G141308	<b>NIST No.</b>	747V66-240502-00		
<b>Standard ID</b>	+0.01000"	<b>Uncertainty</b>		<b>Units</b>		<b>Type</b>	V
<b>Limited Use?</b>	No	<b>Minimum</b>	0.00950	<b>Nominal</b>	0.01000	<b>Maximum</b>	0.01050
<b>Ref Type</b>	Mic Trac	<b>As Found</b>	0.00998	<b>Accuracy</b>	-0.00002	<b>Fail Before</b>	No
		<b>As Left</b>	0.00998	<b>Accuracy</b>	-0.00002	<b>Fail After</b>	No
<b>Gage ID of Standard</b>	REF-G141308						
<b>Std Due Date</b>	10/1/2026	<b>Gage S/N</b>	G141308	<b>NIST No.</b>	747V66-240502-00		
<b>Standard ID</b>	-0.00500"	<b>Uncertainty</b>		<b>Units</b>		<b>Type</b>	V
<b>Limited Use?</b>	No	<b>Minimum</b>	-0.00550	<b>Nominal</b>	-0.00500	<b>Maximum</b>	-0.00450

Ref Type	Mic Trac	As Found	-0.00512	Accuracy	-0.00012	Fail Before	No
		As Left	-0.00512	Accuracy	-0.00012	Fail After	No

Gage ID of Standard REF-G141308  
 Std Due Date 10/1/2026

Gage S/N G141308 NIST No. 747V66-240502-00

Standard ID	-0.01000"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	-0.01050	Nominal	-0.01000	Maximum	-0.00950
Ref Type	Mic Trac	As Found	-0.01011	Accuracy	-0.00011	Fail Before	No
		As Left	-0.01011	Accuracy	-0.00011	Fail After	No

Gage ID of Standard REF-G141308  
 Std Due Date 10/1/2026

Gage S/N G141308 NIST No. 747V66-240502-00

Calibrated By Taichi Daimo Signature  Date: 1/12/2026

**APPROVED**  
 By Kayla Myers at 5:13 pm, Jan 12, 2026