

1115 Garber Road
Broussard LA 70518

Company Peak NDT Solutions	Doc #: 185058	Print#:
Address 515A Verot School Road	OEM #:	Serial#: 13398-2
Lafayette LA	Contact	Location Calibration Lab
Zip: 70506	Dept WO# 062426-07	Mfg.: Lonestar
Page 1 of 1	Cal Date 06/25/2026	Model Profile Gauge
	PO#:	Gage TP-23-50, 4 TPI, 2 TPF, V.038

Comments

Procedure# SDM-26 Rev.1
Accuracy +/- .0002"
Reference Standard Serial#: A111186
Profile Overlay NIST# 03-7-1857

We certify the equipment used for this calibration is traceable to NIST through one or more of the following numbers

Reference Standard Serial#: 8169A
Optical Comparator w/ Optical Edge Detection (16"), Starrett Model HE400
Traceable to NIST via NAMAS certification no: Cert# 17211 Last / Next Cal Dates: 08/15/2023 -> 08/15/2026

Gage Status PASS

Next Calibration Due: 06/25/2029

Certified By: Mason Pellerin

Signature 

This certificate is not valid unless all 1 page(s) are present

NOTE IF ONLY AN "AS FOUND READINGS STATED" AS FOUND AND AS LEFT" READINGS ARE THE SAME WITH NO ADJUSTMENTS

- SDM Calibration certifies the instrument above conforms to the original manufacturer's specifications and has been calibrated using standards whose accuracies are traceable to the (SI) via the National Institute of Standards and Technology (NIST) or have been derived from accepted values of natural physical constants or by the ratio type of self calibration techniques. The uncertainty of measurement has not been applied as part of the decision rules for Pass/Fail status. Calibrations are performed to the latest revision of applicable standards/specifications unless otherwise specified by the customer.
Environmental Conditions (68 Degrees F +/- 2 Degrees) (Humidity 20-55%)

APPROVED

By Kayla Myers at 2:10 pm, Jul 06, 2026

SDM CALIBRATIONS LLC

1115 Garber Road
Broussard LA 70518

Calibration Report

Company	Peak NDT Solutions	Doc #:	185054	Print#:	
Address	515A Verot School Road Lafayette LA	OEM #:		Serial#:	13398-3
Zip:	70506	Contact		Location	Calibration Lab
Page	1 of 1	Dept	WO# 062426-07	Mfg.:	Lonestar
		Cal Date	06/25/2026	Model	Profile Gauge
		PO#:		Gage	TP-542, 4 TPI, 2 TPFV.050

Comments

Procedure# SDM-26 Rev.1
Accuracy +/- .0002"
Reference Standard Serial#: A111157
Profile Overlay NIST# 03-7-1857

We certify the equipment used for this calibration is traceable to NIST through one or more of the following numbers:

Reference Standard Serial#: 8169A
Optical Comparator w/ Optical Edge Detection (16"), Starrett Model HE400
Traceable to NIST via NAMAS certification no: Cert# 17211 Last / Next Cal Dates 08/15/2023 -> 08/15/2026

Gage Status PASS

Next Calibration Due: 06/25/2029

Certified By: Mason Pellerin

Signature 

This certificate is not valid unless all 1 page(s) are present

NOTE IF ONLY AN "AS FOUND READING IS STATED" AS FOUND AND AS LEFT" READINGS ARE THE SAME WITH NO ADJUSTMENTS

- SDM Calibrations certifies the instrument above conforms to the original manufacturer's specifications and has been calibrated using standards whose accuracies are traceable to the (SI) via the National Institute of Standards and Technology (NIST) or have been derived from accepted values of natural physical constants or by the ratio type of self-calibration techniques. The uncertainty of measurement has not been applied as part of the decision rules for Pass/Fail status. Calibrations are performed to the latest revision of applicable standards/specifications unless otherwise specified by the customer.
Environmental Conditions (68 Degrees F +/- 2 Degrees) (Humidity 20-55%)

APPROVED

By Kayla Myers at 2:08 pm, Jul 06, 2026

SDM CALIBRATION\$ LLC

1115 Garber Road
Broussard LA 70518

Calibration Report

Company	Peak NDT Solutions	Doc #:	185055	Print#:	
Address	515A Verot School Road Lafayette LA	OEM #:		Serial#:	13398-4
Zip:	70506	Contact		Location	Calibration Lab
Page	1 of 1	Dept	WO# 062426-07	Mfg.:	Lonestar
		Cal.Date	06/25/2026	Model	Profile Gauge
		PO#:		Gage	TP-56-77, 4 TPI, 3 TPFV.038

Comments

Procedure# SDM-26 Rev.1
Accuracy +/- .0002"
Reference Standard Serial#: A111185
Profile Overlay NIST# 03-7-1857

We certify the equipment used for this calibration is traceable to NIST through one or more of the following numbers

Reference Standard Serial#: 8169A
Optical Comparator w/ Optical Edge Detection (16"), Starrett Model HE400
Traceable to NIST via NAMAS certification no: Cert# 17211 Last / Next Cal Dates: 08/15/2023 -> 08/15/2026

Gage Status PASS

Next Calibration Due: 06/25/2029

Certified By: Mason Pellerin

Signature 

This certificate is not valid unless all 1 page(s) are present

NOTE IF ONLY AN "AS FOUND" READING IS STATED, "AS FOUND AND AS LEFT" READINGS ARE THE SAME WITH NO ADJUSTMENTS

- SDM Calibrations certifies the instrument above conforms to the original manufacturer's specifications and has been calibrated using standards whose accuracies are traceable to the (SI) via the National Institute of Standards and Technology (NIST) or have been derived from accepted values of natural physical constants or by the ratio type of self-calibration techniques. The uncertainty of measurement has not been applied as part of the decision rules for Pass/Fail status. Calibrations are performed to the latest revision of applicable standards/specifications unless otherwise specified by the customer.
Environmental Conditions (68 Degrees F +/- 2 Degrees, Humidity 20-55%)

APPROVED

By Kayla Myers at 2:08 pm, Jul 06, 2026

SDM CALIBRATIONS LLC

1115 Garber Road
Broussard LA 70518

Calibration Report

Company	Peak NDT Solutions	Doc #:	185056	Print#:	
Address	515A Verot School Road Lafayette LA	OEM #:		Serial#:	13398-5
Zip:	70506	Contact		Location	Calibration Lab
Page	1 of 1	Dept	WO# 062426-07	Mfg.:	Lonestar
		Cal.Date	06/25/2026	Model	Profile Gauge
		PO#:		Gage	TP-543, 4 TPI, 3 TPFV.050

Comments

Procedure# SDM-26 Rev.1
Accuracy +/- .0002"
Reference Standard Serial#: A111158
Profile Overlay NIST# 03-7-1857

We certify the equipment used for this calibration is traceable to NIST through one or more of the following numbers

Reference Standard Serial#: 8169A
Optical Comparator w/ Optical Edge Detection (16"), Starrett Model HE400
Traceable to NIST via NAMA certification no: Cert# 17211 Last / Next Cal Dates 08/15/2023 -> 08/15/2026

Gage Status PASS

Next Calibration Due: 06/25/2029

Certified By: Mason Pellerin

Signature 

This certificate is not valid unless all 1 page(s) are present

NOTE IF ONLY AN "AS FOUND READING IS STATED" AS FOUND AND AS LEFT" READINGS ARE THE SAME WITH NO ADJUSTMENTS

- SDM Calibrations certifies the instrument above conforms to the original manufacturer's specifications and has been calibrated using standards whose accuracies are traceable to the (SI) via the National Institute of Standards and Technology (NIST) or have been derived from accepted values of natural physical constants or by the ratio type of self-calibration techniques. The uncertainty of measurement has not been applied as part of the decision rules for Pass/Fail status. Calibrations are performed to the latest revision of applicable standards/specifications unless otherwise specified by the customer.

Environmental Conditions (68 Degrees F +/- 2 Degrees) (Humidity 20-55%)

APPROVED

By Kayla Myers at 2:09 pm, Jul 06, 2026

Company	Peak NDT Solutions	Doc #:	185057	Print#:	
Address	515A Verot School Road Lafayette LA	OEM #:		Serial#:	13398-6
Zip:	70506	Contact		Location	Calibration Lab
Page	1 of 1	Dept	WO# 062426-07	Mfg.:	Lonestar
		Cal Date:	06/25/2026	Model	Profile Gauge
		PO#:		Gage	TP-453, 5 TPI, 3 TPFV.040

Comments

Procedure# SDM-26 Rev.1
 Accuracy +/- .0002"
 Reference Standard Serial#: A111159
 Profile Overlay NIST# 03-7-1857

We certify the equipment used for this calibration is traceable to NIST through one or more of the following numbers

Reference Standard Serial#: 8169A
 Optical Comparator/ Optical Edge Detection (16"), Starrett Model HE400
 Traceable to NIST via NAMA certification no: Cert# 17211 Last / Next Cal Dates 08/15/2023 -> 08/15/2026

Gage Status PASS**Next Calibration Due: 06/25/2029**

Certified By: Mason Pellerin

Signature 

This certificate is not valid unless all 1 page(s) are present

NOTE IF ONLY AN "AS FOUND" READING IS STATED "AS FOUND AND AS LEFT" READINGS ARE THE SAME WITH NO ADJUSTMENTS

- SDM Calibration certifies the instrument above conforms to the original manufacturer's specifications and has been calibrated using standards whose accuracies are traceable to the (SI) via the National Institute of Standards and Technology (NIST) or have been derived from accepted values of natural physical constants or by the ratio type of self-calibration techniques. The uncertainty of measurement has not been applied as part of the decision rules for Pass/Fail status. Calibrations are performed to the latest revision of applicable standards/specifications unless otherwise specified by the customer.
 Environmental Conditions (68 Degrees F +/- 2 Degrees) (Humidity 20-55%)

APPROVED**By Kayla Myers at 2:10 pm, Jul 06, 2026**