

SDM CALIBRATIONS LLC

1115 Garber Road
Broussard LA 70518

Calibration Report

| | | | | | |
|----------------|--|----------------|---------------|-----------------|-----------------|
| Company | Peak NDT Solutions | Doc #: | 185069 | Print#: | |
| Address | 515A Verot School Road Lafayette LA | OEM #: | | Serial#: | L02249 |
| Zip: | 70506 | Contact | | Location | Calibration Lab |
| Page | 1 of 1 | Dept | WO# 062426-07 | Mfg.: | N/A |
| | | CalDate | 06/25/2026 | Model | Step Block |
| | | PO#: | | Gage | .100-.500" |

Step Block Readings As Found

| Master | Actual | Deviation | Serial | Tol +0.001000 / -0.001000 |
|-----------|-----------|-----------|--------|---------------------------|
| +0.100000 | +0.100430 | +0.000430 | | |
| +0.200000 | +0.200090 | +0.000090 | | |
| +0.300000 | +0.300560 | +0.000560 | | |
| +0.400000 | +0.400450 | +0.000450 | | |
| +0.500000 | +0.499590 | -0.000410 | | |

Comments

Procedure# SDM-46 Rev.1
Accuracy +/- .001" per ASTM E797

We certify the equipment used for this calibration is traceable to NIST through one or more of the following numbers

Reference Standard Serial#: K10567

MicTrac (48"), Gagemaker, Model MT4048

Traceable to NIST via NAMAS certification no: Cert# 176480

Last / Next Cal Dates 12/05/2025 -> 12/05/2026

Gage Status PASS

Next Calibration Due: 06/25/2027

Certified By: Mason Pellerin

Signature 

This certificate is not valid unless all 1 page(s) are present

NOTE IF ONLY AN "AS FOUND READING IS STATED" AS FOUND AND AS LEFT" READINGS ARE THE SAME WITH NO ADJUSTMENTS

- SDM Calibrations certifies the instrument above conforms to the original manufacturer's specifications and has been calibrated using standards whose accuracies are traceable to the (SI) via the National Institute of Standards and Technology (NIST) or have been derived from accepted values of natural physical constants or by the ratio type of self-calibration techniques. The uncertainty of measurement has not been applied as part of the decision rules for Pass/Fail status. Calibrations are performed to the latest revisions of applicable standards/specifications unless otherwise specified by the customer.

Environmental Conditions (68 Degrees F +/- 2 Degrees) (Humidity 20-55%)

APPROVED

By Kayla Myers at 2:18 pm, Jul 06, 2026