

SDM CALIBRATION\$ LLC

1115 Garber Road
Broussard LA 70518

Calibration Report

Company	Peak NDT Solutions	Doc #:	185065	Print#:	
Address	515A Verot School Road Lafayette LA	OEM #:		Serial#:	23212
Zip:	70506	Contact		Location	Calibration Lab
Page	1 of 1	Dept	WO# 062426-07	Mfg.:	Dakota
		Cal Date	06/25/2026	Model	Ultrasonic Thickness Gauge
		PO#:		Gage	.025-19.999"

Thickness Readings As Found

Master	Actual	Deviation	Serial	Tol
+0.250000	+0.250000	+0.000000		+0.001000 / -0.001000

Thickness Readings As Found

Master	Actual	Deviation	Serial	Tol
+0.500000	+0.500000	+0.000000		+0.001000 / -0.001000

Thickness Readings As Found

Master	Actual	Deviation	Serial	Tol
+0.750000	+0.751000	+0.001000		+0.001000 / -0.001000

Thickness Readings As Found

Master	Actual	Deviation	Serial	Tol
+1.000000	+1.001000	+0.001000		+0.001000 / -0.001000

Comments

Procedure# SDM-15 Rev.1
Accuracy +/- .001"

We certify the equipment used for this calibration is traceable to NIST through one or more of the following numbers

Reference Standard Serial#: 422711

4-Step Test Block TSP, Model 1018 Steel

Traceable to NIST via NAMAS certification no: Cert# 80602

Last / Next Cal Dates 12/21/2023 -> 12/21/2026

Gage Status PASS

Next Calibration Due: 12/25/2026

Certified By: Mason Pellerin

Signature 

This certificate is not valid unless all 1 page(s) are present

NOTE IF ONLY AN "AS FOUND READINGS STATED" "AS FOUND AND AS LEFT" READINGS ARE THE SAME WITH NO ADJUSTMENTS

- SDM Calibrations certifies the instrument above conforms to the original manufacturer's specifications and has been calibrated using standards whose accuracies are traceable to the (SI) via the National Institute of Standards and Technology (NIST) or have been derived from accepted values of natural physical constants or by the ratio type of self calibration techniques. The uncertainty of measurement hasn't been applied as part of the decision rules for Pass/Fail status. Calibrations are performed to the latest revision of applicable standards/specifications unless otherwise specified by the customer.
Environmental Conditions (68 Degrees F +/- 2 Degrees) (Humidity 20-55%)

APPROVED

By Kayla Myers at 2:17 pm, Jul 06, 2026