



Tubular Inspection Products

1551 Windway
Odessa, TX 79763
(432) 363-2200

PO Box 842
Odessa, TX 79760
Fax: (432) 363-2281

CERTIFICATE OF CALIBRATION

Traceable to NIST — ISO/IEC 17025:2017

This Calibration Certificate confirms that the instruments described herein have been calibrated in compliance with the requirements set forth by ISO/IEC 17025:2017, ensuring that all measurement results are traceable to national or international standards, including those established by the National Institute of Standards and Technology (NIST). This certificate reflects adherence to the manufacturer's specifications, demonstrating compliance with established guidelines for performance and reliability.

Standards: ISO 17025:2017 • ISO 9001 • API Q1/7-2

INSTRUMENT & CALIBRATION DETAILS

Customer	Peak NDT Solutions, LLC	Manufacturer	General
Calibration Date	August 01, 2026	TIP Procedure	TIP-WLMTW01.1
Calibration Due Date	February 01, 2027	Calibration Technician	Ryan Britton
Serial Number	210701029	Order Number	81126
Instrument	White Light Meter	Temperature / Humidity	68.5°F / RH 40.5%

CALIBRATION RESULTS

WL meter was calibrated against a Calibrated Master WL Meter and environment light source. Tolerance: $\pm 5\%$

Calibrated Meter Reading	Actual Reading	Deviation
41.5 fc	41.0 fc	0.5 fc

INSTRUMENT STATUS SUMMARY

Instrument Received	Instrument Returned	Instrument Status
In Tolerance	In Tolerance	PASSED

CALIBRATION STANDARDS USED (TRACEABILITY)

- TIP Master Combo Light Meter, Spectroline XDS-1000 UV / XDS-1000WL / XR-1000, S/N 2036288 / 2036288 / 2036286 (Cert. Due 01/28/2027)
- Magnaflux White Light Meter, S/N 120902340 (Cert. Due 01/28/2027)

Calibration Technician

Ryan Britton

Ryan Britton

Issue Date: August 01, 2026

APPROVED

By Kayla Myers at 2:46 pm, Sep 03, 2026